

TYPICAL PERFORMANCE OF JANIS MODEL HE-3-SSV-PT-2 CRYOGEN-FREE HELIUM-3 CRYOSTAT

Second regenerator heat exchanger was not being used during the condensing;
otherwise the holding time was 115 hours with 3.5 pound copper sample installed.

